

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
 Data File : VX031036.D
 Acq On : 01 Sep 2022 21:42
 Operator : JC/MD
 Sample : N4297-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW9-MW01D1-20220821

Quant Time: Sep 02 02:56:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	221152	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	356542	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	324770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	147170	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	160902	56.545	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.100%	
35) Dibromofluoromethane	5.397	113	152750	58.707	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 117.420%	
50) Toluene-d8	8.653	98	552503	57.697	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 115.400%#	
62) 4-Bromofluorobenzene	11.085	95	190603	56.354	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.700%	
Target Compounds						
17) Carbon Disulfide	2.508	76	2346	0.352	ug/l	99
29) Tetrahydrofuran	5.019	42	9443	8.637	ug/l #	79
44) Trichloroethene	7.123	130	3935	1.188	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
Data File : VX031036.D
Acq On : 01 Sep 2022 21:42
Operator : JC/MD
Sample : N4297-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW9-MW01D1-20220821

Quant Time: Sep 02 02:56:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

