

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
 Data File : VX031041.D
 Acq On : 01 Sep 2022 23:39
 Operator : JC/MD
 Sample : N4297-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D1-20220821

Quant Time: Sep 02 02:57:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	215853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	344113	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	330523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	156883	56.486	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 112.980%	
35) Dibromofluoromethane	5.391	113	152736	60.821	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 121.640%	
50) Toluene-d8	8.653	98	545094	58.979	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 117.960%#	
62) 4-Bromofluorobenzene	11.079	95	184186	56.424	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.840%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.167	85	597	0.269	ug/l	85
9) 1,1,2-Trichlorotrifluo...	2.325	101	6076	2.340	ug/l	91
12) 1,1-Dichloroethene	2.319	96	1927	0.768	ug/l	90
19) Methyl tert-butyl Ether	3.123	73	7136	0.924	ug/l	93
27) cis-1,2-Dichloroethene	4.501	96	9294	2.972	ug/l	88
30) Chloroform	5.105	83	4773	0.963	ug/l #	80
44) Trichloroethene	7.129	130	608096	190.241	ug/l	99
64) Tetrachloroethene	9.275	164	34640	10.497	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
Data File : VX031041.D
Acq On : 01 Sep 2022 23:39
Operator : JC/MD
Sample : N4297-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D1-20220821

Quant Time: Sep 02 02:57:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

