

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
 Data File : VX031043.D
 Acq On : 02 Sep 2022 00:26
 Operator : JC/MD
 Sample : N4297-24
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP08-20220821

Quant Time: Sep 02 02:57:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Pentafluorobenzene		5.556	168	217543	50.000	ug/l	0.00
34) 1,4-Difluorobenzene		6.769	114	357519	50.000	ug/l	0.00
63) Chlorobenzene-d5		10.055	117	321369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4		12.024	152	144462	50.000	ug/l	0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4		5.958	65	156813	56.022	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.040%	
35) Dibromofluoromethane		5.391	113	150514	57.689	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	115.380%	
50) Toluene-d8		8.653	98	548284	57.099	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	114.200%#	
62) 4-Bromofluorobenzene		11.079	95	188768	55.659	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.320%	
Target Compounds							Qvalue
29) Tetrahydrofuran		5.013	42	26592	24.725	ug/l #	83
44) Trichloroethene		7.135	130	1864	0.561	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
Data File : VX031043.D
Acq On : 02 Sep 2022 00:26
Operator : JC/MD
Sample : N4297-24
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP08-20220821

Quant Time: Sep 02 02:57:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

