

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
 Data File : VX031044.D
 Acq On : 02 Sep 2022 00:50
 Operator : JC/MD
 Sample : N4298-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D1-20220822

Quant Time: Sep 02 02:57:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	215522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	364522	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	318128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143751	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	155339	56.016	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.040%
35) Dibromofluoromethane	5.391	113	149035	56.025	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.040%
50) Toluene-d8	8.653	98	545124	55.680	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.360%
62) 4-Bromofluorobenzene	11.085	95	183947	53.196	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.400%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	2473	0.954	ug/l	# 70
27) cis-1,2-Dichloroethene	4.501	96	900	0.288	ug/l	92
44) Trichloroethene	7.129	130	126368	37.320	ug/l	94
64) Tetrachloroethene	9.275	164	6326	1.992	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
Data File : VX031044.D
Acq On : 02 Sep 2022 00:50
Operator : JC/MD
Sample : N4298-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D1-20220822

Quant Time: Sep 02 02:57:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

