



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090122\
Data File : VX031016.D
Acq On : 01 Sep 2022 08:30
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Title : SW846 8260
Last Update : Fri Sep 02 02:49:21 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.1	9011	PASS
75	95	30	60	46.8	27893	PASS
95	95	100	100	100.0	59629	PASS
96	95	5	9	6.5	3889	PASS
173	174	0.00	2	0.6	275	PASS
174	95	50	100	79.0	47083	PASS
175	174	5	9	7.5	3518	PASS
176	174	95	101	95.0	44709	PASS
177	176	5	9	7.6	3390	PASS