

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090123\
 Data File : VX037471.D
 Acq On : 01 Sep 2023 17:44
 Operator : JC/MD
 Sample : 04150-04RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-20230822RE

Quant Time: Sep 02 04:32:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	147281	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	241965	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91399	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87297	41.113	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	82.220%		
35) Dibromofluoromethane	5.385	113	71939	44.892	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.780%		
50) Toluene-d8	8.647	98	280631	48.278	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.560%		
62) 4-Bromofluorobenzene	11.079	95	100044	43.795	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	87.600%		
Target Compounds						
30) Chloroform	5.092	83	4501	1.328	ug/l	96
44) Trichloroethene	7.123	130	3025	1.634	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090123\
Data File : VX037471.D
Acq On : 01 Sep 2023 17:44
Operator : JC/MD
Sample : 04150-04RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-20230822RE

Quant Time: Sep 02 04:32:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

