

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090221\
 Data File : VX024073.D
 Acq On : 02 Sep 2021 16:24
 Operator : JC/MD
 Sample : M3597-03
 Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RO-PSC124042

Quant Time: Sep 02 16:47:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	118122	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	194780	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	171005	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79221	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	82214	49.87	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.74%
35) Dibromofluoromethane	5.397	113	61587	47.77	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.54%
50) Toluene-d8	8.653	98	232644	47.41	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.82%
62) 4-Bromofluorobenzene	11.085	95	80582	44.61	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.22%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.959	59	12502	37.50	ug/l #	76
20) Methylene Chloride	2.782	84	3476	2.47	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090221\
Data File : VX024073.D
Acq On : 02 Sep 2021 16:24
Operator : JC/MD
Sample : M3597-03
Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RO-PSC124042

Quant Time: Sep 02 16:47:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

