

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090221\
 Data File : VX024070.D
 Acq On : 02 Sep 2021 14:13
 Operator : JC/MD
 Sample : VX0902MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0902MBL01

Quant Time: Sep 02 16:11:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	131151	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	225283	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94215	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	102387	55.935	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.880%
35) Dibromofluoromethane	5.391	113	78909	52.920	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.840%
50) Toluene-d8	8.653	98	285872	50.372	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.740%
62) 4-Bromofluorobenzene	11.085	95	103770	49.670	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090221\
Data File : VX024070.D
Acq On : 02 Sep 2021 14:13
Operator : JC/MD
Sample : VX0902MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0902MBL01

Quant Time: Sep 02 16:11:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

