

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090221\
 Data File : VX024075.D
 Acq On : 02 Sep 2021 17:11
 Operator : JC/MD
 Sample : M3620-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-TB-20210830

Quant Time: Sep 03 05:39:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	128098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	217820	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85068	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	90141	50.419	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.840%
35) Dibromofluoromethane	5.397	113	69011	47.868	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.740%
50) Toluene-d8	8.653	98	253303	46.163	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.320%
62) 4-Bromofluorobenzene	11.085	95	90705	44.904	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090221\
Data File : VX024075.D
Acq On : 02 Sep 2021 17:11
Operator : JC/MD
Sample : M3620-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-TB-20210830

Quant Time: Sep 03 05:39:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

