

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090221\
 Data File : VX024079.D
 Acq On : 02 Sep 2021 18:45
 Operator : JC/MD
 Sample : M3620-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-680-682

Quant Time: Sep 03 05:40:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	134017	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	222589	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211004	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88186	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	91807	49.083	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.160%		
35) Dibromofluoromethane	5.397	113	71354	48.433	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.860%		
50) Toluene-d8	8.653	98	263185	46.936	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	93.880%		
62) 4-Bromofluorobenzene	11.085	95	93480	45.286	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.580%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.173	85	956	0.676	ug/l	89
9) 1,1,2-Trichlorotrifluo...	2.331	101	8805	6.545	ug/l	95
12) 1,1-Dichloroethene	2.325	96	5784	4.497	ug/l	95
16) Acetone	2.386	43	4413	5.277	ug/l #	82
17) Carbon Disulfide	2.520	76	1097	0.358	ug/l #	95
27) cis-1,2-Dichloroethene	4.507	96	2067	1.264	ug/l	93
30) Chloroform	5.111	83	5465	1.964	ug/l	99
32) 1,1,1-Trichloroethane	5.404	97	1100	0.462	ug/l #	49
44) Trichloroethene	7.135	130	149336	88.691	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	1366	0.840	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090221\
Data File : VX024079.D
Acq On : 02 Sep 2021 18:45
Operator : JC/MD
Sample : M3620-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-680-682

Quant Time: Sep 03 05:40:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

