

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090221\
 Data File : VX024085.D
 Acq On : 02 Sep 2021 21:30
 Operator : JC/MD
 Sample : M3620-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-640-642

Quant Time: Sep 03 05:42:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	127467	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	213126	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202701	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	87667	49.278	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.560%
35) Dibromofluoromethane	5.397	113	68921	48.858	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.720%
50) Toluene-d8	8.653	98	248896	46.359	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.720%
62) 4-Bromofluorobenzene	11.085	95	87429	44.235	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.480%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.173	85	688	0.512	ug/l	77
9) 1,1,2-Trichlorotrifluo...	2.331	101	3883	3.035	ug/l	97
12) 1,1-Dichloroethene	2.325	96	3480	2.845	ug/l	96
16) Acetone	2.392	43	5589	7.026	ug/l	97
17) Carbon Disulfide	2.514	76	2207	0.756	ug/l #	92
27) cis-1,2-Dichloroethene	4.495	96	2243	1.443	ug/l	93
30) Chloroform	5.105	83	5161	1.951	ug/l	91
44) Trichloroethene	7.135	130	202403	125.545	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	1895	1.217	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090221\
Data File : VX024085.D
Acq On : 02 Sep 2021 21:30
Operator : JC/MD
Sample : M3620-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-640-642

Quant Time: Sep 03 05:42:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

