

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090223\
 Data File : VX037481.D
 Acq On : 02 Sep 2023 14:28
 Operator : JC/MD
 Sample : 04270-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-033123

Quant Time: Sep 04 04:27:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	158929	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	263641	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225713	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94993	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94012	41.031	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.060%
35) Dibromofluoromethane	5.385	113	79166	45.341	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.680%
50) Toluene-d8	8.647	98	303636	47.941	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.880%
62) 4-Bromofluorobenzene	11.079	95	104751	42.086	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.180%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	5011	5.590	ug/l	98
20) Methylene Chloride	2.794	84	3059	1.526	ug/l #	80
51) 4-Methyl-2-Pentanone	8.580	43	7671	3.140	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090223\
Data File : VX037481.D
Acq On : 02 Sep 2023 14:28
Operator : JC/MD
Sample : 04270-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-033123

Quant Time: Sep 04 04:27:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

