

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090223\
 Data File : VX037486.D
 Acq On : 02 Sep 2023 16:27
 Operator : JC/MD
 Sample : 04269-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14B-45.5-083123

Quant Time: Sep 04 04:29:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	169881	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	283608	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	248509	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105030	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	101076	41.270	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.540%
35) Dibromofluoromethane	5.385	113	86089	45.834	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.660%
50) Toluene-d8	8.647	98	332038	48.735	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.460%
62) 4-Bromofluorobenzene	11.079	95	115664	43.199	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.400%
Target Compounds						
44) Trichloroethene	7.129	130	1995	0.919	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090223\
Data File : VX037486.D
Acq On : 02 Sep 2023 16:27
Operator : JC/MD
Sample : 04269-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14B-45.5-083123

Quant Time: Sep 04 04:29:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

