

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090223\
 Data File : VX037491.D
 Acq On : 02 Sep 2023 18:27
 Operator : JC/MD
 Sample : 04266-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-04B-26.5-083123-FD

Quant Time: Sep 04 04:31:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	150629	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	249513	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95397	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90601	41.721	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.440%
35) Dibromofluoromethane	5.385	113	75472	45.672	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.340%
50) Toluene-d8	8.647	98	291239	48.587	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.180%
62) 4-Bromofluorobenzene	11.079	95	103280	43.844	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.680%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.495	96	2690	1.295	ug/l	72
44) Trichloroethene	7.128	130	8032	4.207	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090223\
Data File : VX037491.D
Acq On : 02 Sep 2023 18:27
Operator : JC/MD
Sample : 04266-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-04B-26.5-083123-FD

Quant Time: Sep 04 04:31:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

