

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090223\
 Data File : VX037493.D
 Acq On : 02 Sep 2023 19:14
 Operator : JC/MD
 Sample : 04272-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-01B-67-083123

Quant Time: Sep 04 04:32:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	158126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	262107	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102984	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94203	41.323	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.640%
35) Dibromofluoromethane	5.385	113	78529	45.239	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.480%
50) Toluene-d8	8.647	98	304187	48.309	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.620%
62) 4-Bromofluorobenzene	11.079	95	109170	44.118	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.240%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	422	0.252	ug/l	95
12) 1,1-Dichloroethene	2.319	96	600	0.369	ug/l #	62
27) cis-1,2-Dichloroethene	4.501	96	1747	0.801	ug/l	77
44) Trichloroethene	7.129	130	8685	4.330	ug/l	94
64) Tetrachloroethene	9.275	164	822	0.485	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090223\
Data File : VX037493.D
Acq On : 02 Sep 2023 19:14
Operator : JC/MD
Sample : 04272-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-01B-67-083123

Quant Time: Sep 04 04:32:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

