

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090223\
 Data File : VX037494.D
 Acq On : 02 Sep 2023 19:38
 Operator : JC/MD
 Sample : 04256-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Groundwater

Quant Time: Sep 04 04:32:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	154059	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	258207	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225258	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96761	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92135	41.483	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.960%
35) Dibromofluoromethane	5.385	113	77954	45.586	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.180%
50) Toluene-d8	8.647	98	297323	47.932	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.860%
62) 4-Bromofluorobenzene	11.079	95	105391	43.234	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.460%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2305	2.653	ug/l	95
20) Methylene Chloride	2.794	84	1910	0.983	ug/l #	81
27) cis-1,2-Dichloroethene	4.489	96	742	0.349	ug/l	64
31) Cyclohexane	5.458	56	1560	0.625	ug/l	89
39) Methylcyclohexane	7.373	83	1552	0.608	ug/l	89
44) Trichloroethene	7.129	130	3066	1.552	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090223\
 Data File : VX037494.D
 Acq On : 02 Sep 2023 19:38
 Operator : JC/MD
 Sample : 04256-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Groundwater

Quant Time: Sep 04 04:32:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

