

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090223\
 Data File : VX037497.D
 Acq On : 02 Sep 2023 20:50
 Operator : JC/MD
 Sample : 04273-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1815

Quant Time: Sep 04 04:34:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	155336	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	255210	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219184	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91014	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92315	41.222	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.440%
35) Dibromofluoromethane	5.385	113	77931	46.108	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.220%
50) Toluene-d8	8.647	98	295563	48.208	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.420%
62) 4-Bromofluorobenzene	11.079	95	100436	41.685	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.380%
Target Compounds						
44) Trichloroethene	7.141	130	1210	0.620	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090223\
Data File : VX037497.D
Acq On : 02 Sep 2023 20:50
Operator : JC/MD
Sample : 04273-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1815

Quant Time: Sep 04 04:34:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

