

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090225\
 Data File : VX047572.D
 Acq On : 02 Sep 2025 10:22
 Operator : JC/MD
 Sample : VX0902MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0902MBL01

Quant Time: Sep 03 02:12:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 29 02:31:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	179585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	357090	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	358182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	179432	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	147723	53.566	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.140%
35) Dibromofluoromethane	5.397	113	125503	51.208	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.420%
50) Toluene-d8	8.647	98	445730	51.886	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.780%
62) 4-Bromofluorobenzene	11.079	95	180082	55.800	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090225\
Data File : VX047572.D
Acq On : 02 Sep 2025 10:22
Operator : JC/MD
Sample : VX0902MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0902MBL01

Quant Time: Sep 03 02:12:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 29 02:31:18 2025
Response via : Initial Calibration

