

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090225\
 Data File : VX047573.D
 Acq On : 02 Sep 2025 10:43
 Operator : JC/MD
 Sample : VX0902WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0902WBL01

Quant Time: Sep 03 02:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 29 02:31:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	329307	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	667815	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	660687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	325255	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	272217	53.830	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.660%
35) Dibromofluoromethane	5.397	113	232593	50.746	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.500%
50) Toluene-d8	8.653	98	819224	50.992	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.980%
62) 4-Bromofluorobenzene	11.079	95	338775	56.130	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090225\
Data File : VX047573.D
Acq On : 02 Sep 2025 10:43
Operator : JC/MD
Sample : VX0902WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0902WBL01

Quant Time: Sep 03 02:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 29 02:31:18 2025
Response via : Initial Calibration

