

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090225\
 Data File : VX047581.D
 Acq On : 02 Sep 2025 14:05
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 03 02:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 29 02:31:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	266443	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	536190	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	529066	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	266266	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	219179	53.568	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.140%	
35) Dibromofluoromethane	5.397	113	186859	50.776	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.560%	
50) Toluene-d8	8.653	98	656198	50.871	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.740%	
62) 4-Bromofluorobenzene	11.079	95	273878	56.517	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.040%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090225\
Data File : VX047581.D
Acq On : 02 Sep 2025 14:05
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 03 02:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 29 02:31:18 2025
Response via : Initial Calibration

