



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090225\
Data File : VX047570.D
Acq On : 02 Sep 2025 08:42
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
Title : SW846 8260
Last Update : Fri Aug 29 02:31:18 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.6	21807	PASS
75	95	30	60	50.4	59213	PASS
95	95	100	100	100.0	117421	PASS
96	95	5	9	6.3	7343	PASS
173	174	0.00	2	1.2	1056	PASS
174	95	50	100	73.0	85685	PASS
175	174	5	9	7.4	6310	PASS
176	174	95	101	95.5	81827	PASS
177	176	5	9	6.7	5444	PASS