

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090319\
 Data File : VX012039.D
 Acq On : 03 Sep 2019 23:10
 Operator : SY/SP
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA X/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 04 05:28:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.461	0.452	2.0	92	0.00
3 P	Chloromethane	0.472	0.513	-8.7	99	0.00
4 C	Vinyl Chloride	0.444	0.478	-7.7#	97	0.00
5 T	Bromomethane	0.220	0.269	-22.3#	112	0.00
6 T	Chloroethane	0.276	0.306	-10.9	101	0.00
7 T	Trichlorofluoromethane	0.690	0.706	-2.3	93	0.00
8 T	Diethyl Ether	0.238	0.268	-12.6	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.511	0.511	0.0	90	0.00
10 T	Methyl Iodide	0.414	0.380	8.2	73	0.00
11 T	Tert butyl alcohol	0.066	0.064	3.0	94	0.00
12 CM	1,1-Dichloroethene	0.492	0.518	-5.3#	94	0.00
13 T	Acrolein	0.021	0.014	33.3#	74	0.00
14 T	Allyl chloride	0.891	0.964	-8.2	96	0.00
15 T	Acrylonitrile	0.150	0.156	-4.0	95	0.00
16 T	Acetone	0.165	0.142	13.9	78	0.00
17 T	Carbon Disulfide	1.439	1.616	-12.3	93	0.00
18 T	Methyl Acetate	0.388	0.480	-23.7#	119	0.00
19 T	Methyl tert-butyl Ether	1.603	1.743	-8.7	99	0.00
20 T	Methylene Chloride	0.662	0.668	-0.9	98	0.00
21 T	trans-1,2-Dichloroethene	0.578	0.629	-8.8	96	0.00
22 T	Diisopropyl ether	1.816	1.984	-9.3	100	0.00
23 T	Vinyl Acetate	1.175	1.207	-2.7	90	0.00
24 P	1,1-Dichloroethane	1.053	1.132	-7.5	98	0.00
25 T	2-Butanone	0.210	0.212	-1.0	90	0.00
26 T	2,2-Dichloropropane	0.972	0.898	7.6	85	0.00
27 T	cis-1,2-Dichloroethene	0.677	0.731	-8.0	98	0.00
28 T	Bromochloromethane	0.420	0.420	0.0	87	0.00
29 T	Tetrahydrofuran	0.126	0.135	-7.1	97	0.00
30 C	Chloroform	1.130	1.219	-7.9#	99	0.00
31 T	Cyclohexane	0.846	0.852	-0.7	92	0.00
32 T	1,1,1-Trichloroethane	0.984	1.058	-7.5	97	0.00
33 S	1,2-Dichloroethane-d4	0.690	0.739	-7.1	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.307	0.315	-2.6	92	0.00
36 T	1,1-Dichloropropene	0.440	0.442	-0.5	93	0.00
37 T	Ethyl Acetate	0.239	0.243	-1.7	93	0.00
38 T	Carbon Tetrachloride	0.464	0.469	-1.1	96	0.00
39 T	Methylcyclohexane	0.511	0.491	3.9	91	0.00
40 TM	Benzene	1.264	1.341	-6.1	98	0.00
41 T	Methacrylonitrile	0.143	0.149	-4.2	98	0.00
42 TM	1,2-Dichloroethane	0.455	0.486	-6.8	100	0.00
43 T	Isopropyl Acetate	0.487	0.505	-3.7	97	0.00
44 TM	Trichloroethene	0.341	0.355	-4.1	97	0.00
45 C	1,2-Dichloropropane	0.327	0.343	-4.9#	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090319\
 Data File : VX012039.D
 Acq On : 03 Sep 2019 23:10
 Operator : SY/SP
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA X/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 04 05:28:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.191	0.200	-4.7	99	0.00
47 T	Bromodichloromethane	0.485	0.512	-5.6	98	0.00
48 T	Methyl methacrylate	0.245	0.257	-4.9	97	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	94	0.00
50 S	Toluene-d8	1.128	1.189	-5.4	93	0.00
51 T	4-Methyl-2-Pentanone	0.248	0.255	-2.8	97	0.00
52 CM	Toluene	0.823	0.867	-5.3#	98	0.00
53 T	t-1,3-Dichloropropene	0.502	0.504	-0.4	94	0.00
54 T	cis-1,3-Dichloropropene	0.558	0.569	-2.0	96	0.00
55 T	1,1,2-Trichloroethane	0.273	0.274	-0.4	94	0.00
56 T	Ethyl methacrylate	0.390	0.408	-4.6	98	0.00
57 T	1,3-Dichloropropane	0.470	0.491	-4.5	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.176	0.177	-0.6	90	0.00
59 T	2-Hexanone	0.181	0.180	0.6	94	0.00
60 T	Dibromochloromethane	0.325	0.342	-5.2	100	0.00
61 T	1,2-Dibromoethane	0.263	0.278	-5.7	99	0.00
62 S	4-Bromofluorobenzene	0.525	0.509	3.0	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.312	0.345	-10.6	105	0.00
65 PM	Chlorobenzene	0.987	1.019	-3.2	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.380	-5.3	101	0.00
67 C	Ethyl Benzene	1.798	1.850	-2.9#	97	0.00
68 T	m/p-Xylenes	0.663	0.678	-2.3	96	0.00
69 T	o-Xylene	0.642	0.667	-3.9	98	0.00
70 T	Styrene	1.116	1.165	-4.4	97	0.00
71 P	Bromoform	0.198	0.206	-4.0	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.669	3.866	-5.4	97	0.00
74 T	N-amyl acetate	1.085	1.105	-1.8	94	0.00
75 P	1,1,2,2-Tetrachloroethane	0.765	0.783	-2.4	96	0.00
76 T	1,2,3-Trichloropropane	0.611	0.702	-14.9	109	0.00
77 T	Bromobenzene	0.820	0.889	-8.4	99	0.00
78 T	n-propylbenzene	4.432	4.612	-4.1	96	0.00
79 T	2-Chlorotoluene	2.685	2.794	-4.1	98	0.00
80 T	1,3,5-Trimethylbenzene	3.215	3.342	-4.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.258	0.239	7.4	87	0.00
82 T	4-Chlorotoluene	3.200	3.346	-4.6	95	0.00
83 T	tert-Butylbenzene	3.064	3.108	-1.4	95	0.00
84 T	1,2,4-Trimethylbenzene	3.304	3.459	-4.7	97	0.00
85 T	sec-Butylbenzene	3.806	3.850	-1.2	94	0.00
86 T	p-Isopropyltoluene	3.461	3.484	-0.7	94	0.00
87 T	1,3-Dichlorobenzene	1.644	1.713	-4.2	99	0.00
88 T	1,4-Dichlorobenzene	1.644	1.651	-0.4	94	0.00
89 T	n-Butylbenzene	3.362	3.321	1.2	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090319\
Data File : VX012039.D
Acq On : 03 Sep 2019 23:10
Operator : SY/SP
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA X/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 04 05:28:53 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.691	0.677	2.0	94	0.00
91 T	1,2-Dichlorobenzene	1.481	1.537	-3.8	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.160	0.157	1.9	90	0.00
93 T	1,2,4-Trichlorobenzene	1.054	1.000	5.1	89	0.00
94 T	Hexachlorobutadiene	0.531	0.495	6.8	88	0.00
95 T	Naphthalene	2.281	2.264	0.7	91	0.00
96 T	1,2,3-Trichlorobenzene	0.927	0.897	3.2	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6