

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090320\
 Data File : VX018226.D
 Acq On : 03 Sep 2020 11:59
 Operator : JC/SP
 Sample : PB131388TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131388TB

Quant Time: Sep 04 01:57:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	481625	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	784309	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	772417	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	345180	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	348106	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
35) Dibromofluoromethane	5.47	113	268955	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
50) Toluene-d8	8.70	98	970997	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
62) 4-Bromofluorobenzene	11.12	95	357531	45.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.36%	
Target Compounds						
16) Acetone	2.43	43	31825	12.215	ug/l #	87
18) Methyl Acetate	2.75	43	7448	1.185	ug/l #	75
20) Methylene Chloride	2.84	84	12964	1.296	ug/l	94
43) Isopropyl Acetate	6.41	43	122413	9.020	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090320\
Data File : VX018226.D
Acq On : 03 Sep 2020 11:59
Operator : JC/SP
Sample : PB131388TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131388TB

Quant Time: Sep 04 01:57:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

