

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090320\
 Data File : VX018232.D
 Acq On : 03 Sep 2020 14:14
 Operator : JC/SP
 Sample : L3903-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ140I-20200902

Quant Time: Sep 04 02:26:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	485283	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	773476	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	738116	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	345312	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	345227	51.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.48%	
35) Dibromofluoromethane	5.47	113	269889	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
50) Toluene-d8	8.70	98	942565	48.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.08%	
62) 4-Bromofluorobenzene	11.12	95	339889	44.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.06%	
Target Compounds						
16) Acetone	2.43	43	7200	2.743	ug/l #	86
24) 1,1-Dichloroethane	3.68	63	20435	2.135	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090320\
Data File : VX018232.D
Acq On : 03 Sep 2020 14:14
Operator : JC/SP
Sample : L3903-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ140I-20200902

Quant Time: Sep 04 02:26:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

