

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090320\
 Data File : VX018234.D
 Acq On : 03 Sep 2020 14:59
 Operator : JC/SP
 Sample : L3903-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ118I-20200902

Quant Time: Sep 04 02:31:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	469509	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	784818	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	720959	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	341762	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	343419	52.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.38%	
35) Dibromofluoromethane	5.46	113	276854	51.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.60%	
50) Toluene-d8	8.70	98	944913	47.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.92%	
62) 4-Bromofluorobenzene	11.12	95	339491	43.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.68%	
Target Compounds						
3) Chloromethane	1.31	50	6611	1.097	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090320\
Data File : VX018234.D
Acq On : 03 Sep 2020 14:59
Operator : JC/SP
Sample : L3903-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ118I-20200902

Quant Time: Sep 04 02:31:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

