

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090320\
 Data File : VX018239.D
 Acq On : 03 Sep 2020 16:52
 Operator : JC/SP
 Sample : L3903-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ122D-20200902

Quant Time: Sep 04 02:46:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	427790	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	714462	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	675100	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	331173	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	324798	54.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.38%	
35) Dibromofluoromethane	5.47	113	254772	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
50) Toluene-d8	8.70	98	870243	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
62) 4-Bromofluorobenzene	11.12	95	322303	45.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.40%	
Target Compounds						
24) 1,1-Dichloroethane	3.67	63	9939	1.178	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090320\
Data File : VX018239.D
Acq On : 03 Sep 2020 16:52
Operator : JC/SP
Sample : L3903-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ122D-20200902

Quant Time: Sep 04 02:46:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

