

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090320\
 Data File : VX018241.D
 Acq On : 03 Sep 2020 17:38
 Operator : JC/SP
 Sample : L3903-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-DUP02-20200902

Quant Time: Sep 04 02:57:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	431135	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	706224	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	673302	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	331949	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	323375	54.03	ug/l	0.00
Spiked Amount						
			Recovery	=	108.06%	
35) Dibromofluoromethane	5.46	113	252411	52.48	ug/l	0.00
Spiked Amount						
			Recovery	=	104.96%	
50) Toluene-d8	8.70	98	854246	47.69	ug/l	0.00
Spiked Amount						
			Recovery	=	95.38%	
62) 4-Bromofluorobenzene	11.12	95	311296	44.17	ug/l	0.00
Spiked Amount						
			Recovery	=	88.34%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090320\
Data File : VX018241.D
Acq On : 03 Sep 2020 17:38
Operator : JC/SP
Sample : L3903-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-DUP02-20200902

Quant Time: Sep 04 02:57:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

