

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090320\
 Data File : VX018246.D
 Acq On : 03 Sep 2020 19:32
 Operator : JC/SP
 Sample : L3876-14 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1576

Quant Time: Sep 04 06:55:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	441024	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	708468	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	695812	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	350848	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	316925	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
35) Dibromofluoromethane	5.46	113	252764	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
50) Toluene-d8	8.70	98	861595	47.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.88%	
62) 4-Bromofluorobenzene	11.12	95	339983	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
Target Compounds						
11) Tert butyl alcohol	3.00	59	8430	7.081	ug/l #	94
16) Acetone	2.42	43	11012	4.616	ug/l	89
95) Naphthalene	13.82	128	346905	14.952	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090320\
Data File : VX018246.D
Acq On : 03 Sep 2020 19:32
Operator : JC/SP
Sample : L3876-14 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1576

Quant Time: Sep 04 06:55:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

