

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
 Data File : VX024095.D
 Acq On : 03 Sep 2021 12:20
 Operator : JC/MD
 Sample : M3620-04DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-660-662DL

Quant Time: Sep 03 13:45:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	145934	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	241332	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97563	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109875	53.946	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 107.900%		
35) Dibromofluoromethane	5.397	113	84833	53.110	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.220%		
50) Toluene-d8	8.653	98	312998	51.484	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.960%		
62) 4-Bromofluorobenzene	11.085	95	112194	50.131	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 100.260%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	3233	2.207	ug/l	94
10) Methyl Iodide	2.465	142	1789	0.877	ug/l	93
12) 1,1-Dichloroethene	2.325	96	2831	2.021	ug/l	98
16) Acetone	2.386	43	2005	2.202	ug/l #	75
30) Chloroform	5.099	83	1403	0.463	ug/l #	81
44) Trichloroethene	7.129	130	81634	44.717	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
Data File : VX024095.D
Acq On : 03 Sep 2021 12:20
Operator : JC/MD
Sample : M3620-04DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-660-662DL

Quant Time: Sep 03 13:45:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

