

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
 Data File : VX024096.D
 Acq On : 03 Sep 2021 12:44
 Operator : JC/MD
 Sample : M3620-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-EB-20210831

Quant Time: Sep 03 13:46:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	147153	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	242290	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228023	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93955	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	110125	53.620	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.240%	
35) Dibromofluoromethane	5.397	113	84613	52.763	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.520%	
50) Toluene-d8	8.653	98	307887	50.443	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.880%	
62) 4-Bromofluorobenzene	11.085	95	108649	48.355	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.720%	
Target Compounds						Qvalue
11) Tert butyl alcohol	2.971	59	19133	46.066	ug/l	99
16) Acetone	2.392	43	21157	23.039	ug/l	95
95) Naphthalene	13.780	128	378	1.336	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
Data File : VX024096.D
Acq On : 03 Sep 2021 12:44
Operator : JC/MD
Sample : M3620-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-EB-20210831

Quant Time: Sep 03 13:46:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

