

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
 Data File : VX024102.D
 Acq On : 03 Sep 2021 15:05
 Operator : JC/MD
 Sample : M3643-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OF-3

Quant Time: Sep 04 01:29:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132525	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	220888	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207429	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90375	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99872	53.996	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.000%
35) Dibromofluoromethane	5.397	113	76879	52.585	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.160%
50) Toluene-d8	8.653	98	278338	50.021	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.040%
62) 4-Bromofluorobenzene	11.085	95	100617	49.119	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.240%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	12113	14.647	ug/l #	85
25) 2-Butanone	4.574	43	2237	1.786	ug/l #	68
51) 4-Methyl-2-Pentanone	8.573	43	4323	1.700	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
Data File : VX024102.D
Acq On : 03 Sep 2021 15:05
Operator : JC/MD
Sample : M3643-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OF-3

Quant Time: Sep 04 01:29:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

