

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
 Data File : VX024105.D
 Acq On : 03 Sep 2021 16:16
 Operator : JC/MD
 Sample : M3653-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 04 01:30:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	134077	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	221480	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100763	53.847	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.700%	
35) Dibromofluoromethane	5.397	113	78247	53.377	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.760%	
50) Toluene-d8	8.653	98	279911	50.169	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.340%	
62) 4-Bromofluorobenzene	11.085	95	96983	47.219	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.440%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
Data File : VX024105.D
Acq On : 03 Sep 2021 16:16
Operator : JC/MD
Sample : M3653-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 04 01:30:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

