

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
 Data File : VX024106.D
 Acq On : 03 Sep 2021 16:40
 Operator : JC/MD
 Sample : M3653-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Sep 04 01:30:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136215	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	228668	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213364	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91139	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102460	53.894	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.780%	
35) Dibromofluoromethane	5.397	113	80172	52.971	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.940%	
50) Toluene-d8	8.653	98	288696	50.117	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.240%	
62) 4-Bromofluorobenzene	11.085	95	103110	48.624	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.240%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
Data File : VX024106.D
Acq On : 03 Sep 2021 16:40
Operator : JC/MD
Sample : M3653-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Sep 04 01:30:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

