

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
 Data File : VX024107.D
 Acq On : 03 Sep 2021 17:03
 Operator : JC/MD
 Sample : M3653-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Sep 04 01:31:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	137974	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	228873	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	212840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90437	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105321	54.693	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.380%
35) Dibromofluoromethane	5.397	113	82606	54.531	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.060%
50) Toluene-d8	8.653	98	298598	51.789	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.580%
62) 4-Bromofluorobenzene	11.085	95	105721	49.810	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
Data File : VX024107.D
Acq On : 03 Sep 2021 17:03
Operator : JC/MD
Sample : M3653-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Sep 04 01:31:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

