

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
 Data File : VX024108.D
 Acq On : 03 Sep 2021 17:27
 Operator : JC/MD
 Sample : M3643-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OF-2

Quant Time: Sep 04 01:31:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	125785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	210776	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73734	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	95979	54.671	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.340%	
35) Dibromofluoromethane	5.397	113	74220	53.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.400%	
50) Toluene-d8	8.653	98	268286	50.527	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.060%	
62) 4-Bromofluorobenzene	11.085	95	87641	44.837	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.680%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	15998	20.381	ug/l	91
25) 2-Butanone	4.587	43	2643	2.223	ug/l	98
51) 4-Methyl-2-Pentanone	8.580	43	7925	3.267	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
Data File : VX024108.D
Acq On : 03 Sep 2021 17:27
Operator : JC/MD
Sample : M3643-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OF-2

Quant Time: Sep 04 01:31:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

