

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
 Data File : VX024112.D
 Acq On : 03 Sep 2021 19:01
 Operator : JC/MD
 Sample : M3646-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 99-01

Quant Time: Sep 04 01:33:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	135771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	227121	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209130	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96957	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103116	54.417	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.840%	
35) Dibromofluoromethane	5.403	113	78950	52.519	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.040%	
50) Toluene-d8	8.653	98	281696	49.235	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.460%	
62) 4-Bromofluorobenzene	11.085	95	105014	49.859	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.720%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.971	59	17241	44.990	ug/l	97
16) Acetone	2.386	43	63007	74.364	ug/l	88
20) Methylene Chloride	2.794	84	932	0.576	ug/l #	92
25) 2-Butanone	4.574	43	35632	27.765	ug/l	99
51) 4-Methyl-2-Pentanone	8.580	43	7707	2.948	ug/l	88
52) Toluene	8.726	92	1133	0.281	ug/l #	75
59) 2-Hexanone	9.433	43	24581	12.233	ug/l	84
68) m/p-Xylenes	10.305	106	1136	0.391	ug/l	96
69) o-Xylene	10.646	106	1261	0.442	ug/l	94
80) 1,3,5-Trimethylbenzene	11.457	105	1801	0.296	ug/l	86
84) 1,2,4-Trimethylbenzene	11.756	105	3273	0.540	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
Data File : VX024112.D
Acq On : 03 Sep 2021 19:01
Operator : JC/MD
Sample : M3646-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
99-01

Quant Time: Sep 04 01:33:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

