

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
 Data File : VX024090.D
 Acq On : 03 Sep 2021 08:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 03 13:42:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	74	0.00
2 T	Dichlorodifluoromethane	0.528	0.506	4.2	65	0.00
3 P	Chloromethane	0.579	0.504	13.0	61	0.00
4 C	Vinyl Chloride	0.565	0.514	9.0#	62	0.00
5 T	Bromomethane	0.355	0.331	6.8	67	0.00
6 T	Chloroethane	0.383	0.337	12.0	64	0.00
7 T	Trichlorofluoromethane	0.892	0.920	-3.1	71	0.00
8 T	Diethyl Ether	0.311	0.310	0.3	68	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.502	0.542	-8.0	74	0.00
10 T	Methyl Iodide	0.699	0.683	2.3	64	0.00
11 T	Tert butyl alcohol	0.141	0.127	9.9	62	0.00
12 CM	1,1-Dichloroethene	0.480	0.490	-2.1#	71	0.00
13 T	Acrolein	0.084	0.100	-19.0	83	0.00
14 T	Allyl chloride	0.884	0.881	0.3	69	0.00
15 T	Acrylonitrile	0.314	0.291	7.3	64	0.00
16 T	Acetone	0.312	0.350	-12.2	81	0.00
17 T	Carbon Disulfide	1.144	1.127	1.5	66	0.00
18 T	Methyl Acetate	0.779	0.721	7.4	66	0.00
19 T	Methyl tert-butyl Ether	1.670	1.830	-9.6	75	0.00
20 T	Methylene Chloride	0.596	0.570	4.4	70	0.00
21 T	trans-1,2-Dichloroethene	0.513	0.530	-3.3	72	0.00
22 T	Diisopropyl ether	1.761	1.827	-3.7	71	0.00
23 T	Vinyl Acetate	1.425	1.484	-4.1	70	0.00
24 P	1,1-Dichloroethane	0.980	1.030	-5.1	73	0.00
25 T	2-Butanone	0.473	0.466	1.5	67	0.00
26 T	2,2-Dichloropropane	0.805	0.948	-17.8	81	0.00
27 T	cis-1,2-Dichloroethene	0.610	0.633	-3.8	72	0.00
28 T	Bromochloromethane	0.446	0.445	0.2	68	0.00
29 T	Tetrahydrofuran	0.306	0.278	9.2	62	0.00
30 C	Chloroform	1.038	1.115	-7.4#	75	0.00
31 T	Cyclohexane	0.897	0.902	-0.6	69	0.00
32 T	1,1,1-Trichloroethane	0.889	1.012	-13.8	77	0.00
33 S	1,2-Dichloroethane-d4	0.698	0.708	-1.4	73	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	71	0.00
35 S	Dibromofluoromethane	0.331	0.355	-7.3	73	0.00
36 T	1,1-Dichloropropene	0.463	0.514	-11.0	72	0.00
37 T	Ethyl Acetate	0.569	0.557	2.1	66	0.00
38 T	Carbon Tetrachloride	0.483	0.580	-20.1	77	0.00
39 T	Methylcyclohexane	0.564	0.630	-11.7	71	0.00
40 TM	Benzene	1.369	1.467	-7.2	71	0.00
41 T	Methacrylonitrile	0.309	0.309	0.0	66	0.00
42 TM	1,2-Dichloroethane	0.538	0.621	-15.4	76	0.00
43 T	Isopropyl Acetate	0.890	0.930	-4.5	69	0.00
44 TM	Trichloroethene	0.378	0.414	-9.5	73	0.00
45 C	1,2-Dichloropropane	0.364	0.384	-5.5#	69	0.00
46 T	Dibromomethane	0.253	0.283	-11.9	73	0.00
47 T	Bromodichloromethane	0.458	0.549	-19.9	76	0.00
48 T	Methyl methacrylate	0.434	0.457	-5.3	67	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
 Data File : VX024090.D
 Acq On : 03 Sep 2021 08:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 03 13:42:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	61	0.00
50 S	Toluene-d8	1.260	1.261	-0.1	68	0.00
51 T	4-Methyl-2-Pentanone	0.576	0.592	-2.8	65	0.00
52 CM	Toluene	0.887	0.977	-10.1#	72	0.00
53 T	t-1,3-Dichloropropene	0.521	0.628	-20.5	75	0.00
54 T	cis-1,3-Dichloropropene	0.555	0.644	-16.0	74	0.00
55 T	1,1,2-Trichloroethane	0.365	0.407	-11.5	73	0.00
56 T	Ethyl methacrylate	0.569	0.630	-10.7	70	0.00
57 T	1,3-Dichloropropane	0.606	0.661	-9.1	70	0.00
58 T	2-Chloroethyl Vinyl ether	0.294	0.287	2.4	64	0.00
59 T	2-Hexanone	0.442	0.485	-9.7	68	0.00
60 T	Dibromochloromethane	0.353	0.427	-21.0	74	0.00
61 T	1,2-Dibromoethane	0.386	0.427	-10.6	70	0.00
62 S	4-Bromofluorobenzene	0.464	0.514	-10.8	73	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	71	0.00
64 T	Tetrachloroethene	0.399	0.447	-12.0	75	0.00
65 PM	Chlorobenzene	1.043	1.153	-10.5	73	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.443	-19.4	76	0.00
67 C	Ethyl Benzene	1.822	2.078	-14.1#	73	0.00
68 T	m/p-Xylenes	0.694	0.812	-17.0	74	0.00
69 T	o-Xylene	0.682	0.791	-16.0	74	0.00
70 T	Styrene	1.111	1.354	-21.9	76	0.00
71 P	Bromoform	0.263	0.341	-29.7#	78	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.814	3.973	-4.2	76	0.00
74 T	N-amyl acetate	1.724	1.729	-0.3	72	0.00
75 P	1,1,2,2-Tetrachloroethane	1.250	1.215	2.8	75	0.00
76 T	1,2,3-Trichloropropane	1.166	1.165	0.1	74	0.00
77 T	Bromobenzene	0.939	0.974	-3.7	76	0.00
78 T	n-propylbenzene	4.362	4.648	-6.6	76	0.00
79 T	2-Chlorotoluene	2.636	2.795	-6.0	77	0.00
80 T	1,3,5-Trimethylbenzene	3.134	3.419	-9.1	77	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.407	-2.5	73	0.00
82 T	4-Chlorotoluene	2.993	3.274	-9.4	78	0.00
83 T	tert-Butylbenzene	3.044	3.341	-9.8	78	0.00
84 T	1,2,4-Trimethylbenzene	3.125	3.418	-9.4	78	0.00
85 T	sec-Butylbenzene	3.854	4.309	-11.8	78	0.00
86 T	p-Isopropyltoluene	3.172	3.653	-15.2	79	0.00
87 T	1,3-Dichlorobenzene	1.699	1.842	-8.4	80	0.00
88 T	1,4-Dichlorobenzene	1.695	1.833	-8.1	79	0.00
89 T	n-Butylbenzene	2.832	3.222	-13.8	79	0.00
90 T	Hexachloroethane	0.488	0.585	-19.9	80	0.00
91 T	1,2-Dichlorobenzene	1.646	1.754	-6.6	79	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.262	-1.9	71	0.00
93 T	1,2,4-Trichlorobenzene	0.991	1.103	-11.3	80	0.00
94 T	Hexachlorobutadiene	0.442	0.524	-18.6	89	0.00
95 T	Naphthalene	3.231	3.526	-9.1	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090321\
Data File : VX024090.D
Acq On : 03 Sep 2021 08:59
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 03 13:42:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.987	1.115	-13.0	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6