

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031087.D
 Acq On : 03 Sep 2022 17:03
 Operator : JC/MD
 Sample : N4357-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D2-20220823

Quant Time: Sep 06 00:14:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	230953	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	359908	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	317865	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139468	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	138511	46.611	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.220%		
35) Dibromofluoromethane	5.391	113	134573	51.237	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	102.480%		
50) Toluene-d8	8.653	98	470106	48.633	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.260%		
62) 4-Bromofluorobenzene	11.085	95	154990	45.396	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.800%		
Target Compounds					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	32928	11.851	ug/l	91
12) 1,1-Dichloroethene	2.319	96	917	0.342	ug/l #	70
27) cis-1,2-Dichloroethene	4.483	96	4193	1.253	ug/l	50
44) Trichloroethene	7.129	130	267288	79.950	ug/l	95
49) 1,4-Dioxane	7.696	88	2031	26.476	ug/l #	80
52) Toluene	8.720	92	6335	0.925	ug/l	97
64) Tetrachloroethene	9.275	164	28043	8.836	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031087.D
Acq On : 03 Sep 2022 17:03
Operator : JC/MD
Sample : N4357-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D2-20220823

Quant Time: Sep 06 00:14:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

