

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031091.D
 Acq On : 03 Sep 2022 18:37
 Operator : JC/MD
 Sample : N4357-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW4-20220823

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/06/2022
 Supervised By : Mahesh Dadoda 09/06/2022

Quant Time: Sep 06 00:15:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	217135	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	340472	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	305265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133064	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	135871	48.632	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.260%		
35) Dibromofluoromethane	5.391	113	130227	52.413	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	104.820%		
50) Toluene-d8	8.653	98	465304	50.884	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.760%		
62) 4-Bromofluorobenzene	11.085	95	153830	47.629	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.260%		
Target Compounds					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	11743	4.496	ug/l	91
12) 1,1-Dichloroethene	2.319	96	2905	1.152	ug/l	92
24) 1,1-Dichloroethane	3.617	63	1519	0.334	ug/l #	82
27) cis-1,2-Dichloroethene	4.489	96	3906	1.242	ug/l	81
30) Chloroform	5.111	83	1825	0.366	ug/l #	80
38) Carbon Tetrachloride	5.684	117	1624	0.450	ug/l #	67
44) Trichloroethene	7.129	130	1906623	602.859	ug/l	99
49) 1,4-Dioxane	7.696	88	836m	11.520	ug/l	
55) 1,1,2-Trichloroethane	9.159	97	919	0.334	ug/l #	62
64) Tetrachloroethene	9.275	164	11661	3.826	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031091.D
Acq On : 03 Sep 2022 18:37
Operator : JC/MD
Sample : N4357-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW4-20220823

Manual Integrations
APPROVED

Reviewed By : John Carlone 09/06/2022
Supervised By : Mahesh Dadoda 09/06/2022

Quant Time: Sep 06 00:15:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

