

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031099.D
 Acq On : 03 Sep 2022 21:44
 Operator : JC/MD
 Sample : N4357-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20220823

Quant Time: Sep 06 00:18:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	213386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	339235	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	303915	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132317	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	136386	49.674	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.340%
35) Dibromofluoromethane	5.397	113	126761	51.204	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.400%
50) Toluene-d8	8.653	98	452462	49.660	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.320%
62) 4-Bromofluorobenzene	11.079	95	151015	46.928	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.860%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	500	0.228	ug/l	88
9) 1,1,2-Trichlorotrifluo...	2.331	101	162946	63.476	ug/l	91
12) 1,1-Dichloroethene	2.325	96	2161	0.872	ug/l #	65
27) cis-1,2-Dichloroethene	4.495	96	3977	1.286	ug/l	80
38) Carbon Tetrachloride	5.666	117	1139	0.316	ug/l #	62
44) Trichloroethene	7.129	130	348617	110.632	ug/l	99
52) Toluene	8.726	92	2086	0.323	ug/l	97
64) Tetrachloroethene	9.275	164	19501	6.427	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031099.D
Acq On : 03 Sep 2022 21:44
Operator : JC/MD
Sample : N4357-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20220823

Quant Time: Sep 06 00:18:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

