

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031100.D
 Acq On : 03 Sep 2022 22:08
 Operator : JC/MD
 Sample : N4357-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP10-20220823

Quant Time: Sep 06 00:18:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/06/2022
 Supervised By : Mahesh Dadoda 09/07/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	212085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	332900	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	300102	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127768	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	132955	48.721	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.440%		
35) Dibromofluoromethane	5.391	113	126358	52.012	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	104.020%		
50) Toluene-d8	8.653	98	437834	48.969	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.940%		
62) 4-Bromofluorobenzene	11.085	95	143120	45.321	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.640%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.319	101	1014m	0.397	ug/l	Qvalue
44) Trichloroethene	7.135	130	9646	3.119	ug/l	85
76) 1,2,3-Trichloropropane	11.238	75	896m	0.297	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031100.D
Acq On : 03 Sep 2022 22:08
Operator : JC/MD
Sample : N4357-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP10-20220823

Quant Time: Sep 06 00:18:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 09/06/2022
Supervised By : Mahesh Dadoda 09/07/2022

