

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031110.D
 Acq On : 04 Sep 2022 02:48
 Operator : JC/MD
 Sample : PB147428TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147428TB

Quant Time: Sep 06 01:29:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	225213	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	353853	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	324859	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144194	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	134450	46.397	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.800%		
35) Dibromofluoromethane	5.391	113	130952	50.711	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.420%		
50) Toluene-d8	8.653	98	467983	49.242	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.480%		
62) 4-Bromofluorobenzene	11.079	95	162571	48.432	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.860%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	18690	16.707	ug/l	98
20) Methylene Chloride	2.794	84	15022	3.480	ug/l #	80
37) Ethyl Acetate	4.721	43	23692	7.539	ug/l #	94
43) Isopropyl Acetate	6.342	43	105288	21.436	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031110.D
Acq On : 04 Sep 2022 02:48
Operator : JC/MD
Sample : PB147428TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147428TB

Quant Time: Sep 06 01:29:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

