

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031111.D
 Acq On : 04 Sep 2022 03:11
 Operator : JC/MD
 Sample : PB147400TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147400TB

Quant Time: Sep 06 01:29:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	220587	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	345941	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	311219	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132222	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135827	47.855	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.720%
35) Dibromofluoromethane	5.385	113	130321	51.621	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.240%
50) Toluene-d8	8.653	98	459274	49.431	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.860%
62) 4-Bromofluorobenzene	11.079	95	156411	47.662	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.320%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	17509	15.980	ug/l	89
20) Methylene Chloride	2.794	84	9393	1.393	ug/l #	80
37) Ethyl Acetate	4.721	43	23641	7.695	ug/l #	93
43) Isopropyl Acetate	6.342	43	105004	21.867	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031111.D
Acq On : 04 Sep 2022 03:11
Operator : JC/MD
Sample : PB147400TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147400TB

Quant Time: Sep 06 01:29:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

