

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031113.D
 Acq On : 04 Sep 2022 03:58
 Operator : JC/MD
 Sample : N4438-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KTS2-WC-S-MH-01-0-20

Quant Time: Sep 06 01:30:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	211218	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	336282	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	305922	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133717	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	139177	51.211	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.420%
35) Dibromofluoromethane	5.391	113	126970	51.739	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.480%
50) Toluene-d8	8.653	98	455709	50.456	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.920%
62) 4-Bromofluorobenzene	11.085	95	155534	48.756	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.520%
Target Compounds						
						Qvalue
16) Acetone	2.391	43	22896	21.823	ug/l	93
20) Methylene Chloride	2.788	84	10182	1.880	ug/l #	87
37) Ethyl Acetate	4.726	43	19718	6.602	ug/l #	94
43) Isopropyl Acetate	6.348	43	91153	19.528	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031113.D
Acq On : 04 Sep 2022 03:58
Operator : JC/MD
Sample : N4438-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KTS2-WC-S-MH-01-0-20

Quant Time: Sep 06 01:30:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

