

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031115.D
 Acq On : 04 Sep 2022 04:45
 Operator : JC/MD
 Sample : N4479-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-8

Quant Time: Sep 06 01:31:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	217578	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	347014	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	314559	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	134693	48.112	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.220%
35) Dibromofluoromethane	5.391	113	131657	51.989	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.980%
50) Toluene-d8	8.653	98	461821	49.551	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.100%
62) 4-Bromofluorobenzene	11.085	95	159718	48.519	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.040%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	21023	19.452	ug/l	99
20) Methylene Chloride	2.788	84	15566	3.899	ug/l #	78
37) Ethyl Acetate	4.721	43	23383	7.587	ug/l #	91
43) Isopropyl Acetate	6.342	43	102240	21.226	ug/l #	91
95) Naphthalene	13.774	128	13828	1.099	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031115.D
Acq On : 04 Sep 2022 04:45
Operator : JC/MD
Sample : N4479-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-8

Quant Time: Sep 06 01:31:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

