

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031117.D
 Acq On : 04 Sep 2022 05:31
 Operator : JC/MD
 Sample : N4479-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-10

Quant Time: Sep 06 01:31:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	212634	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	332272	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	303338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138605	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	130749	47.789	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.580%
35) Dibromofluoromethane	5.391	113	122866	50.670	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.340%
50) Toluene-d8	8.653	98	439971	49.301	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.600%
62) 4-Bromofluorobenzene	11.085	95	153699	48.763	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.520%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	20341	19.258	ug/l	98
20) Methylene Chloride	2.788	84	14976	3.803	ug/l #	86
37) Ethyl Acetate	4.721	43	24108	8.169	ug/l #	93
43) Isopropyl Acetate	6.348	43	101620	22.033	ug/l #	90
95) Naphthalene	13.774	128	29021	2.310	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031117.D
Acq On : 04 Sep 2022 05:31
Operator : JC/MD
Sample : N4479-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-10

Quant Time: Sep 06 01:31:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

