

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031120.D
 Acq On : 04 Sep 2022 06:41
 Operator : JC/MD
 Sample : N4479-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-13

Quant Time: Sep 06 01:33:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	219791	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	340134	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	315803	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	145700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135318	47.849	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.700%
35) Dibromofluoromethane	5.391	113	129940	52.349	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.700%
50) Toluene-d8	8.653	98	458985	50.243	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.480%
62) 4-Bromofluorobenzene	11.079	95	164720	51.051	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.100%
Target Compounds						
					Qvalue	
16) Acetone	2.391	43	19947	18.270	ug/l	99
20) Methylene Chloride	2.788	84	14492	3.414	ug/l	86
37) Ethyl Acetate	4.720	43	24119	7.984	ug/l #	93
43) Isopropyl Acetate	6.354	43	106506	22.559	ug/l #	90
95) Naphthalene	13.774	128	15339	1.161	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031120.D
Acq On : 04 Sep 2022 06:41
Operator : JC/MD
Sample : N4479-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-13

Quant Time: Sep 06 01:33:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

