

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031121.D
 Acq On : 04 Sep 2022 07:05
 Operator : JC/MD
 Sample : N4479-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-14

Quant Time: Sep 06 01:33:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	214862	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	337201	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	307492	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	131162	47.443	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.880%		
35) Dibromofluoromethane	5.391	113	127930	51.988	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	103.980%		
50) Toluene-d8	8.653	98	449922	49.679	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.360%		
62) 4-Bromofluorobenzene	11.079	95	161191	50.392	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.780%		
Target Compounds					Qvalue	
16) Acetone	2.386	43	28125	26.352	ug/l	93
20) Methylene Chloride	2.788	84	14600	3.588	ug/l #	82
25) 2-Butanone	4.574	43	8925	5.417	ug/l	97
37) Ethyl Acetate	4.721	43	24178	8.073	ug/l #	95
43) Isopropyl Acetate	6.348	43	106787	22.815	ug/l #	91
52) Toluene	8.720	92	6969	1.086	ug/l	95
95) Naphthalene	13.774	128	74437	5.953	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031121.D
Acq On : 04 Sep 2022 07:05
Operator : JC/MD
Sample : N4479-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-14

Quant Time: Sep 06 01:33:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

