

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031081.D
 Acq On : 03 Sep 2022 14:29
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 06 00:04:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.514	0.465	9.5	87	0.00
3 P	Chloromethane	0.440	0.375	14.8	87	0.00
4 C	Vinyl Chloride	0.537	0.448	16.6#	79	0.00
5 T	Bromomethane	0.240	0.294	-22.5	114	0.00
6 T	Chloroethane	0.269	0.187	30.5#	63	0.00
7 T	Trichlorofluoromethane	0.908	0.884	2.6	92	0.00
8 T	Diethyl Ether	0.317	0.285	10.1	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.602	0.571	5.1	92	0.00
10 T	Methyl Iodide	0.483	0.772	-59.8#	123	0.00
11 T	Tert butyl alcohol	0.153	0.133	13.1	93	0.01
12 CM	1,1-Dichloroethene	0.581	0.551	5.2#	87	0.00
13 T	Acrolein	0.118	0.066	44.1#	51	0.00
14 T	Allyl chloride	0.771	0.683	11.4	83	0.00
15 T	Acrylonitrile	0.302	0.277	8.3	86	0.00
16 T	Acetone	0.248	0.226	8.9	88	0.00
17 T	Carbon Disulfide	1.509	1.311	13.1	84	0.00
18 T	Methyl Acetate	0.714	0.638	10.6	85	0.00
19 T	Methyl tert-butyl Ether	1.789	1.717	4.0	88	0.00
20 T	Methylene Chloride	0.693	0.597	13.9	87	0.00
21 T	trans-1,2-Dichloroethene	0.646	0.593	8.2	89	0.00
22 T	Diisopropyl ether	1.554	1.410	9.3	83	0.00
23 T	Vinyl Acetate	1.256	1.186	5.6	85	0.00
24 P	1,1-Dichloroethane	1.047	0.965	7.8	85	0.00
25 T	2-Butanone	0.383	0.345	9.9	84	0.00
26 T	2,2-Dichloropropane	0.836	0.719	14.0	82	0.00
27 T	cis-1,2-Dichloroethene	0.724	0.676	6.6	88	0.00
28 T	Bromochloromethane	0.310	0.279	10.0	87	0.00
29 T	Tetrahydrofuran	0.247	0.216	12.6	83	0.00
30 C	Chloroform	1.149	1.076	6.4#	89	0.00
31 T	Cyclohexane	0.879	0.800	9.0	84	0.00
32 T	1,1,1-Trichloroethane	0.985	0.961	2.4	89	0.00
33 S	1,2-Dichloroethane-d4	0.643	0.604	6.1	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.365	0.379	-3.8	96	0.00
36 T	1,1-Dichloropropene	0.511	0.489	4.3	87	0.00
37 T	Ethyl Acetate	0.444	0.421	5.2	82	0.00
38 T	Carbon Tetrachloride	0.531	0.554	-4.3	91	0.00
39 T	Methylcyclohexane	0.633	0.628	0.8	84	0.00
40 TM	Benzene	1.467	1.458	0.6	85	0.00
41 T	Methacrylonitrile	0.237	0.230	3.0	82	0.00
42 TM	1,2-Dichloroethane	0.484	0.494	-2.1	89	0.00
43 T	Isopropyl Acetate	0.694	0.666	4.0	83	0.00
44 TM	Trichloroethene	0.464	0.445	4.1	82	0.00
45 C	1,2-Dichloropropane	0.376	0.360	4.3#	82	0.00
46 T	Dibromomethane	0.282	0.276	2.1	88	0.00
47 T	Bromodichloromethane	0.526	0.529	-0.6	84	0.00
48 T	Methyl methacrylate	0.350	0.345	1.4	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
 Data File : VX031081.D
 Acq On : 03 Sep 2022 14:29
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 06 00:04:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 02:49:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.011	0.010	9.1	84	0.01
50 S	Toluene-d8	1.343	1.333	0.7	87	0.00
51 T	4-Methyl-2-Pentanone	0.446	0.428	4.0	83	0.00
52 CM	Toluene	0.952	0.964	-1.3#	87	0.00
53 T	t-1,3-Dichloropropene	0.521	0.541	-3.8	84	0.00
54 T	cis-1,3-Dichloropropene	0.582	0.588	-1.0	83	0.00
55 T	1,1,2-Trichloroethane	0.405	0.410	-1.2	88	0.00
56 T	Ethyl methacrylate	0.553	0.575	-4.0	85	0.00
57 T	1,3-Dichloropropane	0.640	0.619	3.3	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.276	0.270	2.2	77	-0.01
59 T	2-Hexanone	0.337	0.335	0.6	84	0.00
60 T	Dibromochloromethane	0.420	0.447	-6.4	86	0.00
61 T	1,2-Dibromoethane	0.427	0.448	-4.9	91	0.00
62 S	4-Bromofluorobenzene	0.474	0.478	-0.8	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.499	0.442	11.4	77	0.00
65 PM	Chlorobenzene	1.194	1.181	1.1	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.441	0.467	-5.9	91	0.00
67 C	Ethyl Benzene	2.046	2.064	-0.9#	85	0.00
68 T	m/p-Xylenes	0.802	0.809	-0.9	86	0.00
69 T	o-Xylene	0.787	0.793	-0.8	86	0.00
70 T	Styrene	1.273	1.333	-4.7	86	0.00
71 P	Bromoform	0.343	0.389	-13.4	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	4.031	3.986	1.1	84	0.00
74 T	N-amyl acetate	1.217	1.197	1.6	82	0.00
75 P	1,1,2,2-Tetrachloroethane	1.289	1.250	3.0	89	0.00
76 T	1,2,3-Trichloropropane	1.182	1.183	-0.1	90	0.00
77 T	Bromobenzene	1.039	1.002	3.6	86	0.00
78 T	n-propylbenzene	4.591	4.528	1.4	85	0.00
79 T	2-Chlorotoluene	2.827	2.725	3.6	86	0.00
80 T	1,3,5-Trimethylbenzene	3.392	3.310	2.4	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.376	0.363	3.5	80	0.00
82 T	4-Chlorotoluene	3.163	3.061	3.2	87	0.00
83 T	tert-Butylbenzene	3.376	3.573	-5.8	91	0.00
84 T	1,2,4-Trimethylbenzene	3.371	3.441	-2.1	89	0.00
85 T	sec-Butylbenzene	4.422	4.220	4.6	80	0.00
86 T	p-Isopropyltoluene	3.502	3.632	-3.7	87	0.00
87 T	1,3-Dichlorobenzene	1.991	1.858	6.7	80	0.00
88 T	1,4-Dichlorobenzene	1.982	1.923	3.0	88	0.00
89 T	n-Butylbenzene	3.157	3.194	-1.2	80	0.00
90 T	Hexachloroethane	0.635	0.695	-9.4	92	0.00
91 T	1,2-Dichlorobenzene	2.042	2.032	0.5	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.314	0.346	-10.2	95	0.00
93 T	1,2,4-Trichlorobenzene	1.292	1.488	-15.2	95	0.00
94 T	Hexachlorobutadiene	0.566	0.604	-6.7	96	0.00
95 T	Naphthalene	4.532	4.675	-3.2	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031081.D
Acq On : 03 Sep 2022 14:29
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 06 00:04:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 02:49:21 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.355	1.475	-8.9	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6