



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090322\
Data File : VX031080.D
Acq On : 03 Sep 2022 13:59
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090122W.M
Title : SW846 8260
Last Update : Fri Sep 02 02:49:21 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.1	7872	PASS
75	95	30	60	45.4	23605	PASS
95	95	100	100	100.0	52035	PASS
96	95	5	9	6.8	3525	PASS
173	174	0.00	2	1.3	551	PASS
174	95	50	100	82.0	42653	PASS
175	174	5	9	7.0	3005	PASS
176	174	95	101	95.7	40840	PASS
177	176	5	9	7.0	2871	PASS